

SPRAY DRYER

Facility Requirement Guide

ADL-311SA and GB-210 require:

1-Air compression system

- Recommended Air source requirement: 28L/min and 3Kg/cm² compressed air
- Minimum requirement: 13.5L/min and 2kg/cm² compressor air

2-Electrical power source and male plug

- Recommended plug specification: 250V and 30A

DL-410 requires:

1-Air compression system

- Recommended specifications: 28L/min and 8Kg/cm² compressed air
- Minimum requirement: 13.5 L/min and 4Kg/cm² compressed air

2-Electrical power source and male plug

- Recommended plug specification: 250V and 30A

GAS-410 requires:

1-N₂ gas source

- 99% or higher (99.998 recommended)
- Minimum requirement: 13.5L/min and 2kg/cm²

2-Electrical power source and male plug

- Recommended plug specification: 250V and 30A

PLEASE SEE YAMATO RECOMMENDATION FOR COMPRESSOR, AIR COMBINATION, POWER PLUGS.