

SINCE 1889



Sterilizer

Model

SK101C/111C

Instruction Manual

Jan. 17, 2015

- First Edition -

- Thank you for purchasing "Sterilizer, SK Series" of Yamato Scientific Co., Ltd.
- To use this unit properly, read this "Instruction Manual" thoroughly before using this unit. Keep this instruction manual around this unit for referring at anytime.



WARNING!

Carefully read and thoroughly understand the important warning items described in this manual before using this unit.

Yamato Scientific America Inc.

◆ Cautions in Using with Safety	1
• Explanation.....	1
• Table of Illustrated Symbols	2
• Fundamental Matters of "WARNING!" and "CAUTION!"	3
• Fundamental Matters of "WARNING!" and "CAUTION!"	4
◆ Before Using This Unit	5
• Requirements for Installation	5
• Requirements for Installation	6
• Requirements for Installation	7
• Installation Procedure	8
• Installation Procedure	9
◆ Description and Function of Each Part	10
• Main Unit	10
• Control Panel (Keys).....	11
◆ Operation Method	12
• Apparatus/Liquid Sterilization Course	12
• Apparatus/Liquid Sterilization Course	13
• Sterilizing Parameters.....	14
◆ Handling Precautions.....	15
• WARNING	15
• CAUTION	16
◆ Maintenance Method	17
• Daily Inspection and Maintenance	17
◆ Long storage and disposal	18
• When not using this unit for long term / When disposing.....	18
◆ In the Event of Failure.....	19
• Safety Device and Error Code	19
◆ After Service and Warranty	20
◆ Specification	21
◆ Wiring Diagram	22
◆ Replacement Parts Table	23
◆ Reference	24
• List of Dangerous Substances	24

Explanation

MEANING OF ILLUSTRATED SYMBOLS

Illustrated Symbols

Various symbols are used in this safety manual in order to use the unit without danger of injury and damage of the unit. A list of problems caused by ignoring the warnings and improper handling is divided as shown below. Be sure that you understand the warnings and cautions in this manual before operating the unit.



WARNING! If the warning is ignored, there is the danger of a problem that may cause a serious accident or even fatality.



CAUTION! If the caution is ignored, there is the danger of a problem that may cause injury/damage to property or the unit itself.

Meaning of Symbols



This symbol indicates items that urge the warning (including the caution). A detailed warning message is shown adjacent to the symbol.



This symbol indicates items that are strictly prohibited. A detailed message is shown adjacent to the symbol with specific actions not to perform.



This symbol indicates items that should be always performed. A detailed message with instructions is shown adjacent to the symbol.

Table of Illustrated Symbols

Warning



Warning
generally



Warning
high voltage



Warning
high temperature



Warning
drive train



Warning
explosive

Caution



Caution
generally



Caution
electrical shock



Caution
scald



Caution
no road heating



Caution
not to drench



Caution
water only



Caution
deadly poison

Prohibit



Prohibit
generally



Prohibit
inflammable



Prohibit
to disassemble



Prohibit
to touch

Compulsion



Compulsion
generally



Compulsion
connect to the
grounding terminal



Compulsion
install on a flat
surface



Compulsion
disconnect the power
plug



Compulsion
periodical
inspection

Cautions in Using with Safety

Fundamental Matters of "WARNING!" and "CAUTION!"

WARNING!

Do not use this unit in an area where there is flammable or explosive gas

Never use this unit in an area where there is flammable or explosive gas. This unit is not explosion-proof. An arc may be generated when the power switch is turned on or off, and fire/explosion may result. (Refer to page 24 "List of Dangerous Substances".)

Always ground this unit

Always ground this unit on the power equipment side in order to avoid electrical shock due to a power surge.

If a problem occurs

If smoke or strange odor should come out of this unit for some reason, turn off the circuit breaker right away, and then disconnect the power plug or power terminal. Immediately contact a service technician for inspection. If this procedure is not followed, fire or electrical shock may result. Never perform repair work yourself, since it is dangerous and not recommended.

Do not use the power cord if it is bundled or tangled

Do not use the power cord if it is bundled or tangled. If it is used in this manner, it could overheat and fire may be caused.

Do not process, bend, wring, or stretch the power cord forcibly

Do not process, bend, wring, or stretch the power cord forcibly. Fire or electrical shock may result.

Substances that can not be used

Never use explosive substances, flammable substances and substances that include explosive or flammable ingredients in this unit. Explosion or fire may occur. (Refer to page 24 "List of Dangerous Substances".)

Do not disassemble or modify this unit

Do not disassemble or modify this unit. Fire or electrical shock or failure may be caused.

Do not get close to the vapor outlet / Do not block the outlet


The vapor outlet is provided on the left face of equipment. Do not put your hands or face close to the outlet. Do not block the outlet. A burn injury or equipment failure may result in.

Fundamental Matters of "WARNING!" and "CAUTION!"

CAUTION!

When opening the cover...

Make sure that the pressure of equipment has decreased to 0(zero) MPa before opening the cover. Generally the cover does not open due to the safety lock mechanism under the high pressure condition. The high-temperature and pressure vapor blows out if the cover is forced open under high pressure, which may cause a burn injury. A large amount of vapor blows out from inside of the chamber when opening the cover just after the sterilizing operation has completed (when the temperature inside the chamber is high). Do not put your hands and face close to the cover.



When draining water...

Make sure that the pressure of equipment has decreased to **0(zero)** MPa before draining the sterilizing water. The hot water blows out if the valve is opened under high pressure. The sterilizing water remains very hot just after the sterilizing operation has completed even the pressure reading is 0(zero) MPa. Drain the water after it is sufficiently cooled down.

Do not touch the hot section

Some sections on the equipment such as the circumference of cover or drain bottle are very hot during or just after the operation of equipment. Do not touch these sections to avoid burn injury.

When taking the sterile samples from the chamber...

Sufficiently remove the vapor inside the chamber before taking the sterile samples from the chamber. Wear heat-resistant leather gloves to take them from the chamber to protect your hands from high-temperature samples.

Do not touch the heat releasing outlet

Do not directly touch the heat releasing outlet located around the outer covering. The vapor may blow out from the safety valve by an accident during sterilizing operation. Do not block the outlet.

During a thunder storm

During a thunderstorm, turn off the power key immediately, then turn off the circuit breaker and the main power. If this procedure is not followed, fire or electrical shock may be caused.

Requirements for Installation


1. Choose a proper place for installation



- Do not install this unit in a place where:
 - ◆ Rough or dirty surface.
 - ◆ Flammable gas or corrosive gas is generated.
 - ◆ Ambient temperature 35°C and above or 5°C and below.
 - ◆ Ambient temperature fluctuates violently.
 - ◆ There is direct sunlight.
 - ◆ There is excessive humidity and dust.
 - ◆ There is a constant vibration.
 - ◆ Not horizontal surface.
 - ◆ The power source is instable.




- Install this unit on a stable place with the space as shown below.



2. Installation on horizontal surface



- Use the equipment on the horizontal and firm place to keep the water inside the chamber horizontal. If the equipment tilts and the heater appears from the water surface, the temperature on the area above the water rises and a heater failure or operation stop due to water level detector function may occur.
- The weight of main unit is approximately 14kg. Carry and install the equipment carefully by two or more persons.



Requirements for Installation

3. Before/after installing



- It may cause injury to a person if this unit falls down or moves by the earthquake and the impact, etc.. To prevent, take safety measures that the unit cannot fall down, and do not install it to a busy place.

4. Ventilate the equipment sufficiently




- Do not block the heat releasing outlets on the side face and back face of equipment during operation. The temperature inside the equipment rises, which may cause the deterioration or failure of equipment, accident, or fire disaster.

5. Do not use this unit in an area where there is flammable or explosive gas



- Never use this unit in an area where there is flammable or explosive gas. This unit is not explosion-proof. An arc may be generated when the power switch is turned ON or OFF, and fire/explosion may result.
- To know about flammable or explosive gas, refer to page 24 "List of Dangerous Substances".



6. Choose a correct power distribution board or receptacle



- Choose a correct power distribution board or receptacle that meets the unit's rated electric capacity.
- Operating voltage range for respective equipment models are as follows.
SK101C models: AC100 to 120V SK111C models: AC200 to 240V

Electric capacity: SK101C: AC100V-120V 12.5A-15A, SK111C: AC200V-240V 7.5A-9A
(NOTE)

There could be a case that the unit won't run even after turning ON the power. Inspect whether the voltage of the main power is lower than the specified value, or whether the other device(s) uses the same power line of this unit. If these phenomena should be found, change the power line of this unit to the other power line.

- Starburst connection with a branching receptacle or extended wiring with a cord reel lowers electrical power voltage, which may cause the degradation of refrigeration capability.
- Connect the unit only to the recommended power supply. If this is connected to a gas pipe, water pipe or telephone line, it may result to an accident or malfunction, could cause injuries or fatality.

Requirements for Installation


7. Handling of power code

- ⊘ Do not entangle the power cord. This will cause overheating and possibly a fire.
- Do not bend or twist the power cord, or apply excessive tension to it. This may cause a fire and electrical shock.
- Do not lay the power cord under a desk or chair, and do not allow it to be pinched in order to prevent it from being damaged and to avoid a fire or electrical shock.
- Keep the power cord away from any heating equipment such as a room heater. The cord's insulation may melt and cause a fire or electrical shock.
- ❗ If the power cord becomes damaged (wiring exposed, breakage, etc.), immediately turn off the power at the rear of this unit and shut off the main supply power. Then contact your nearest dealer for replacement of the power cord. Leaving it may cause a fire or electrical shock.
- Connect the power plug to the receptacle which is supplied appropriate power and voltage.

8. Always ground this unit


- ⊘ Be sure to connect the earth wire (the green cable of power cord) to the grounding conductor or ground terminal to prevent accidents caused by electric leakage.
- ❗ Do not connect the earth wire to gas or water pipes. If not, fire disaster may be caused.
- ⊘ Do not connect the earth wire to the ground for telephone wire or lightning conductor. If not, fire disaster or electric shock may be caused.
- Please consult your local electrical contractor for power connecting work.
- Do not use a branching receptacle, which may cause the heat generation.
- The D class earth connecting works is required if no ground terminal is provided. In this case, consult with the selling office where you purchased or our sales office.
- Securely connect it to the switchboard or outlet.

SK101C



An earth connecting outlet is recommended to be used.

SK111C





These models do not include the power plug. Correctly connect the ground to fit with the power supply facility to be connected.

9. Connect the power cord paying attention to the color of each core wire

- ❗ When connecting the power cord, do check the breaker on the electric power equipment be "OFF". Note that SK101C and SK111C do not equip with the power plug. Select and connect the appropriate plug or terminal corresponding to the power capacity that is adjusted to the status of the power supply equipment side.

Core Wire Color	Interior Wiring
Black	Power Supply Side
White	Power Supply Side
Green	Ground Wire Side


Installation Procedure

<p>1</p>	<p>Set the attached drain board onto the bottom surface inside the chamber</p> <ul style="list-style-type: none"> The drain board stabilizes the sterile samples inside the chamber as well as protects the heater and sensor. Make sure to set it. 	 <p>Set the drain board horizontally.</p> <p>The water level indicator board</p>
<p>2</p>	<p>Close the drain valve</p> <ul style="list-style-type: none"> Close the drain valve at the bottom on the right side face of main unit. Water leak occurs if not fully closed, which may cause the burn injury no-load (water) operation. The drain valve and lead it to the draining site. 	
<p>3</p>	<p>Pour water into the chamber</p> <ul style="list-style-type: none"> Before setting the sterile samples, pour water into the chamber to the water level gauge (notch) position. Insufficient water may cause the no-load (water) operation. Check the water level every time before operation. Refill it before the level becomes too low. Water is required to be poured at dissolution operation, as well as sterilizing operation. When the water level lowers, the equipment cuts off the heater. Depending on the conditions of equipment, however, the detection requires too much time, which may cause the heater deterioration. Refill water before the water level becomes too low. <p>Refer to the right table for the quantity of water to be refilled.</p>	
<p>4</p>	<p>Use distilled or purified water for sterilizing water</p> <ul style="list-style-type: none"> Fill distilled or purified water inside the chamber. Tap water may be used. Calculus generate inside the chamber when tap water is used. Frequent cleaning is therefore required. Do not use well water. It may cause the corrosion or dirt inside the chamber. 	
<p>5</p>	<p>Set the sterile samples</p> <ul style="list-style-type: none"> Set the samples to the chamber, putting them into the attached rack or cast (sold separately). Put the sample or sterilization bag into the chamber so they should not block or cover the sensor inside the chamber, exhaust outlet and end connection to pressure gauge. If they are blocked or covered, the vapor cannot be discharged and the equipment cannot be operated correctly. Do not spill the samples when taking them out or putting them into the chamber. The failure in piping system, bad smell or dirt may result in. In case liquid such as medicinal solution or medium is sterilized, the amount of liquid should be 60% or less of the capacity of container. They may be boiled over if too much quantity is supplied. Widely open the opening of sterilization bag when used. If it is closed, the samples are insufficiently sterilized. 	

Installation Procedure

6 Close the door before operation


- Place the sterilizing drum with properly-piled items into the pot with the positioning mark on the pot lid to be aligned to the center of the lower handle. Press the lid gently by hand and turn it clockwise for the upper and lower handles to close completely and the safety lock to remain in open position, and the safety lock will lock the upper and lower handles automatically when the pressure in pot rises.



- The lock lever on the cover is held by the hook and does not move when the cover remains open.
- Fully close the cover and slide the lock lever on the cover to the left side. If it is closed inappropriately, the vapor blows out from the inside, which may cause a burns or injury to the people around it.
- Do not press the hook and operate the lock lever for purposes other than maintenance of equipment.


Description and Function of Each Part

Main Unit



Description and Function of Each Part


Control Panel (Keys)




No.	Name	Function
1	SET key	This key is used to select temperature/time mode.
2	▲ key	This key is used to increase the temperature/time parameter
3	▼ key	This key is used to decrease the temperature/time parameter
4	Water level lamp flicker	When the sterilizer is short of water, the lamp is flashing, when add water to the "standard" water level, the water level lamp will be on persistently.
5	Press increase lamp flicker	the steam pressure in sterilizer $\geq 0.028\text{Mpa}$, the indicator is light.
6	Interlock lamp flicker	Close the lid, after rotating until closing of the upper and lower handle, the indicator is light.
7	Temperature lamp	The temperature parameter.
8	Time lamp	When set the sterilization time or countdown, the indicator is light.
9	Heating lamp	When the heater is in the working state, the indicator is light, in the process of heat insulation, it is flashing.
10	Sterilizing lamp	When the sterilizer reaches the setting temperature (thermal insulation), the indicator is light.
11	End lamp	after the end of sterilization, the indicator is light (the buzzer warning).
12	Time display	The display sterilizing time

Apparatus/Liquid Sterilization Course


Follow the procedures below for the setting of apparatus sterilization and liquid sterilization courses.




1. Pour the water to the designated scale in chamber, lay the sterilization samples in place and close the cover, as shown in the right figure, the green water level indicator lamp, yellow lid lock indicator lamp and green temp. indicator lamp light up.




2. Press **SET** key to shift the function light to the position of sterilizing temperature, the sterilization temperature indicator on the panel is light, the display shows SP. The sterilizing temperature may be set as required.




3. Press **▲▼** key to raise or reduce the temperature. After adjusting to the required temperature, press again to confirm the new setting and complete the temperature setting procedures.



3. press **SET** key to shift the function lamp to the position of sterilizing time, the sterilization temperature indicator on the panel is light, the display shows ST. The sterilizing time may be set as required.




4. Press **▲** key to increase the time and press **▼** key to reduce it. After adjusting to the required time, press key and **SET** time to confirm the new setting and complete the time setting procedures. The sterilizing time may be set with reference to "Sterilizing Parameters"




5. The running light being on indicates the heater is heating, and with the rise of temperature inside the evaporation pot, the cold air inside the pot will be released from the vent automatically.

Note: The outside surface of this unit may get hot to the touch contact between unprotected skin and outside surface of this unit during operation may cause burns. Always use protective gear to shield operator from hot surfaces.

Apparatus/Liquid Sterilization Course




6. when the temperature rises to above 102 °C, the pressure indicator is light, the top cover cannot be opened




7. When the pressure and temperature reach the set scope, the running light (sterilizing light) will be on and the time will start working for effective heat preservation and sterilization. "Sterilizing Parameters Table" may be referenced for sterilizing time.

Note: After sterilizing, [End] display, the drain valve shall keep open, from which traces of steam will be exhausted to discharge the cold air in the pot, and the sample will be dried and sterilized completely.



8. When countdown for sterilizing finishes, the running light (finish lamp) will be on. After the buzzer rings a few times, the sterilizer will stop working automatically, and the sterilizing operation will complete.



9. Opening the lid: After sterilizing finishes, the power shall be switched off. As the liquor and utensils, etc. may spill or be broken, they shall be cooling off naturally after heating stops, and only after the steam exhaust valve releases the steam till the pressure gauge indicates zero, may the upper handle be turned counterclockwise for the tip of the positioning mark to point to center of the lower handle in order to open the lid.



Note: the lid may only be opened after sterilizing finishes, the remaining steam is released till the pressure gauge indicates zero.

Sterilizing Parameters

Item to be sterilized	Necessary heat preservation time	Relative steam temperature (°C)	Relative steam pressure (MPa)
Rubber	15	121	0.1-0.105
Dressings	30-45	121-126	0.105-0.14
Utensils	15	121-126	0.105-0.14
Apparatus	10	121-126	0.105-0.14
Liquor	20-40	121-126	0.105-0.14
Dishware	20	121-126	0.105-0.14

Use calibration offset function

Calibration offset is a function which corrects the difference between the temperature in furnace and that of controller (sensor temperature) if arises. The function parallel corrects the difference either to the plus or minus side within the whole temperature range of unit. The function can be set or cancelled by the **(SET)** key.



- ① Start operation with the target setting temperature. Check the temperature in furnace (temperature of sample) with a thermograph after it is stabilized.
 - ② Check the difference between the setting temperature and that in furnace (temperature of sample).
 - ③ Press the **(SET)** key. Select the character "SC", which indicates the calibration offset, using the "▲▼", and then press the **(SET)** key.
 - ④ Input the difference using the "▲▼" and then press the **(SET)** key. This completes the setting.
- ❖ The setting range of offset correction temperature is +20°C to plus side and -20°C to minus side respectively.

WARNING



If a problem occurs



If smoke or strange odor should come out of this unit for some reason, turn off the power key right away, and then turn off the circuit breaker and the main power. Immediately contact a service technician for inspection. If this procedure is not followed, fire or electrical shock may result. Never perform repair work yourself, since it is dangerous and not recommended.

Measure for flammability and handling of flammable solvent



This unit is not designed as the explosion-proof construction. Pay special attention to the handling of the sample to be handled with this unit on the consumption with the explosive material, flammable material, and similar ones. The flammable material may be vaporized by leaving it at the temperature higher than room temperature, and could cause the fire or explosion. When handling such material, provide enough ventilation before the operation. (Refer to page 24 "List of Dangerous Substances".)

Do not modify



Do not modify this unit. An accident, malfunction, electric shock or fire may result.

CAUTION

CAUTION!

During a thunder storm



During a thunderstorm, turn off the power key immediately, then turn off the circuit breaker and the main power. If this procedure is not followed, fire or electrical shock may be caused.

Do not touch the hot section



The temperature on the cover and top board on the chamber are very hot during operation or just after operation is completed. Do not touch these sections to avoid a burn injury.

When opening the cover...



Make sure that the pressure gauge reading has decreased to 0(zero) MPa before opening the cover. Open the cover slow carefully. The high-temperature and pressure vapor blows out if the cover is opened during high pressure.

When draining water...



The water in the chamber is very hot just after the operation is completed. Be careful not to get a burn injury. Drain the water after the water is sufficiently cooled down.
Do not drain water during operation. The hot water blows out if the drain valve is opened while the pressure is increasing.

Do not damage the packing on the cover or flange on the chamber



Damage or dirt on these areas may cause the vapor leakage, which may be the cause of burn injury. Keep these sections always clean. Do not damage them with the rack when taking out and putting in the sterile samples. The packing degrades with time. It must be replaced if vapor leak occurs frequently. In this case consult with the selling office where you purchased or our sales office.

Replace the packing early



The packing is a consumable. If it shows the sign of damage or hardening, replace it early. Please consult with the selling office where you purchased or our sales office for the replacement of packing.

Do not perform procedures other than described in this document



Do not perform procedures other than described in this document. Otherwise an unexpected accident may occur.

Daily Inspection and Maintenance

For the safety use of this unit, please perform the daily inspection and maintenance without fail. Using the city water to this unit might attach dirt. Do inspect and maintain this point while performing daily inspection and maintenance.

WARNING!

- Disconnect the power cable from the power source when doing an inspection or maintenance unless needed.
- Perform the daily inspection and maintenance after the temperature of this unit return to its normal state.
- Do not disassemble this unit.

CAUTION!

- Use a well-drained soft cloth to wipe dirt on this unit. Do not use benzene, thinner or cleanser for wiping. Do not scrub this unit. Deformation, deterioration or discoloration may result.

Cleaning inside the chamber

- Use soft sponge to clean inside the chamber not to damage the surface inside the chamber. Do not remove the filter on the bottom of the chamber when cleaning, if removed, the pipe fitting will get clogged with dirt.
- The heater and sensor are provided on the bottom inside the chamber. Make sure not to bend or damage the filter.

For any questions, contact the dealer who you purchased this unit from, or the nearest sales division in our company.

When not using this unit for long term / When disposing

CAUTION!

When not using this unit for long term...

- Turn off the power and disconnect the power cord.

WARNING!

When disposing...

- Keep out of reach of children.
- Consult with the specialized disposal services when disposing the equipment.

Environmental protection should be considered

We request you to disassemble this unit as possible and recycle the reusable parts considering to the environmental protection. The feature components of this unit and materials used are listed below.

Component Name	Material
Exterior Parts	
Outer covering	Bonderizing steel plate baked with melamine resin coating, ABS resin
Chamber, Cover	Stainless steel SUS304
Packing	Silicon rubber
Plates	PET resin film
Electrical Parts	
Switches, Relay	Resin, copper
Circuit boards	Composite of glass fiber and other
Heater	SUS pipe heater
Power cord	Synthetic rubber coated wiring materials, copper and nickel

Safety Device and Error Code

Analysis of Common Failures

Failure	Cause	Solution
1. Not heating after power is on	1. abnormal input power supply 2. heater being damaged 3. incorrect temperature adjustment and time setting	1. to supply the power according to the nameplate 2. to replace the heater 3. to reset after switching off the machine
2. Steam ejecting from the safety valve after temperature rises.	1. the upper and lower handle failing to reach proper positions 2. the safety lock pin being jammed	1. to keep the upper and lower handles contacting properly 2. to move the safety pin
3. Safety valve failing to work after exceeding the parameters	1. the safety valve being blocked 2. the safety valve unable to work properly	1. to unclog the safety valve 2. to replace the safety valve with qualified one
4. The pressure gauge failing to return to zero under non-pressure condition	1. the pressure gauge unable to work properly	1. to replace the pressure gauge with qualified one
5. Steam leaking from between the lid and the pot after pressure rises	1. the seal being not properly placed 2. the seal being damaged	1. to move the seal to a proper position 2. to replace the seal with a qualified one
6. The temperature just rise to 100°C	1. The vent is stuck.	1. to clean the vent and make it smooth 2. If the first solution does not work, please replace the vent with a qualified one

In the case of the error other than listed above occurred, turn off the power switch and primary power source immediately. Contact the shop of your purchase or nearest Yamato Scientific Service Office.

In Case of Request for Repair

If the failure occurs, stop the operation, turn OFF the power switch, and unplug the power plug. Please contact the sales agency that this unit was purchased, or the Yamato Scientific's sales office.

< Check following items before contact >

- ◆ Model Name of Product
 - ◆ Production Number
 - ◆ Purchase Date
 - ◆ About Trouble (in detail as possible)
- } See the production plate attached to this unit.

Minimum Retention Period of Performance Parts for Repair


The minimum retention period of performance parts for repair of this unit is 7 years after discontinuance of this unit.

The "performance part for repair" is the part that is required to maintain this unit.


Product name		Sterilizer	
Model		SK101C	SK111C
Performance	Temperature control range	50to 126°C (sterilization)	
	Maximum Operational Pressure	0.142MPa	
	Operational ambient temperature	5°C~35°C	
Chamber part	Heater	115V /1500W	220V /1500W
	Drain valve	Globe valve	
	liquid level sensor	Float Switch	
Configurations	Temperature control system	PID control by microcomputer	
	Setting/display method	Digital setting by UP/DOWN key / Digital display	
	Timer	Range: 0 to 999min	
	Operation functions	Fixed temperature operation procedure	
Safety devices		Water level detection (liquid expansion method), safety valve (0.142MPa), safety interlocking interactive device, spring full lift safety valve	
Pressure vessel standard		Small-sized pressure vessel (notification of installation is not required)	
Standard	Effective dimensions of chamber (Diameter × Depth mm)	Φ280×H275	
	External dimensions* (Diameter × H mm)	Φ343×H550mm	
	Effective capacity of chamber	18L	
	Weight	17.5kg	
	Power supply (50/60Hz)	AC 115V 15A	AC 220V 9A
Accessories		Rack × 1 SK111A_01_04_02 (dia254×depth240mm)	

*: The external dimension does not include the dimension of projection areas.

SK101C



SK111C



Symbol	Part name	Symbol	Part name
Q1	Power switch	SMJ1	Control board
FU1、2	fuse	SMJ2	Display board
H1	Heater	Pt	Temperature sensor
BL	Folat Switch	SW1	Cover interlock switch

Replacement Parts Table

Common parts

Symbol	Part Name	Code No.	Specification	Manufacturer
Q1	Power switch	A990100037	9-2-1 SK accessories	YSJ
SMJ2	Display board	A020105002	SMJ2	YSJ
	Cable	A011307001		YSJ
Pt	Temperature sensor	A010502008	Pt100Ω	YSJ
BL	Float Switch	A011507001		YSJ
SW1	Cover interlock switch	A011506004		YSJ
	Lid Packing	A081902024		YSJ
	Pressure Gauge	A990100243		YSJ
	Safety valve	A990100249		YSJ
	Steam exhaust valve	A990100255		YSJ
	Rubber Foot	B082099007	C-30-RK-3225	YSJ
	Drain Valve	A990100235		YSJ
T	Terminal block	A011301012	HD320-4PH	YSJ

SK101C

Symbol	Part Name	Code No.	Specification	Manufacturer
SMJ1	CONT board	A020105001	SMJ1 110V	YSJ
H1	Heater	A080501013	1500W/110V	YSJ
FU1、2	Fuse Tube	A010301004	20A 250V	YSJ
	Power Cable	A011208003	3*2.0 With plug	YSJ

SK111C

Symbol	Part Name	Code No.	Specification	Manufacturer
SMJ1	CONT board	A020105003	SMJ1 220V	YSJ
H1	Heater	A080501014	1500W/220V	YSJ
FU1、2	Fuse Tube	A010301002	15A 250V	YSJ
	Power Cable	A011210002	3*2.08 Not plug	YSJ

List of Dangerous Substances



Never use explosive substances, flammable substances and substances that include explosive or flammable ingredients in this unit.

EXPLOSIVE

EXPLOSIVE:	Ethylene glycol dinitrate (nitro glycol), Glycerin trinitrate (nitroglycerine), Cellulose nitrate (nitrocellulose), and other explosive nitrate esters
	Trinitrobenzene, Trinitrotoluene, Trinitrophenol (picric acid), and other explosive nitro compounds
	Acetyl hydroperoxide (peracetic acid), Methyl ethyl ketone peroxide, Benzyl peroxide, and other organic peroxides

FLAMMABLE

IGNITING:	Lithium (metal), Potassium (metal), Sodium (metal), Yellow phosphorus, Phosphorus sulfide, Red phosphorus, Celluloid compounds, Calcium carbide, Lime phosphate, Magnesium (powder), Aluminum (powder), Powder of metals other than magnesium and aluminum, Sodium hydrosulfite
OXIDIZING:	Potassium chlorate, Sodium chlorate, Ammonium chlorate, and other chlorate
	Potassium perchlorate, Sodium perchlorate, Ammonium perchlorate, and other perchlorate
	Potassium peroxide, Sodium peroxide, Barium peroxide, and other inorganic peroxide
	Potassium nitrate, Sodium nitrate, Ammonium nitrate, and other nitrate
	Sodium chlorite and other chlorites
INFLAMMABLE LIQUID:	Calcium hypochlorite and other hypochlorites
	Ethyl ether, Gasoline, Acetaldehyde, Propylene chloride, Carbon disulfide, and other flammable substances having a flash point of lower than -30°C
	Normal hexane, ethylene oxide, acetone, benzene, methyl ethyl ketone, and other flammable substances having a flash point of -30°C or higher but lower than 0°C
	Methanol, Ethanol, Xylene, Pentyl acetate (amyl acetate), and other flammable substances having a flash point of 0°C or higher but lower than 30°C
FLAMMABLE GAS:	Kerosene, Light oil (gas oil), Oil of turpentine, Isopentyl alcohol (isoamyl alcohol), Acetic acid, and other flammable substances having a flash point of 30°C or higher but lower than 65°C
	Hydrogen, Acetylene, Ethylene, Methane, Propane, Butane, and other flammable substances which assume a gaseous state at 15°C and 1 atm

(Source: Appendix Table 1 of Article 6 of the Industrial Safety and Health Order in Japan)

Responsibility

Please follow the instructions in this document when using this unit. Yamato Scientific has no responsibility for the accidents or breakdown of device if it is used with a failure to comply. Never conduct what this document forbids. Unexpected accidents or breakdown may result in.

Note

- ◆ The contents of this document may be changed in future without notice.
- ◆ Any books with missing pages or disorderly binding may be replaced.

**Instruction Manual for
Sterilizer
Model SK101C/111C
First Edition Jan. 17. 2015
Revision**

Yamato Scientific America, Inc.
925 Walsh Ave, Santa Clara, CA 95050
Tel:408-235-7725
For technical information and service
call: 1-800-292-6286
<http://www.yamato-usa.com>